



Approved AHP® Disinfectants for Transducers

TRANSDUCERS

The following transducers have documented approval for the use of .5% AHP products. Approvals are subject to change as manufacturers may modify substrates for manufacturing which affect compatibility. Be advised that this list does not limit the use of .5% AHP products to the listed transducers. If you have questions on about a listed or unlisted product you can reach out to the manufacturer or contact CompatibilityNA@diversey.com

February 2020



Canon/Toshiba

Approved AHP® Disinfectants for Transducers

Canon / Toshiba	PC-20M
Canon / Toshiba	PLI-1205BX
Canon / Toshiba	PLI-2002BT
Canon / Toshiba	PLI-2004BX
Canon / Toshiba	PLI-705BX
Canon / Toshiba	PLT-1005BT
Canon / Toshiba	PLT-1202BT
Canon / Toshiba	PLT-1202S
Canon / Toshiba	PLT-1204BT
Canon / Toshiba	PLT-1204BX
Canon / Toshiba	PLT-1204MV
Canon / Toshiba	PLT-1204ST
Canon / Toshiba	PLT-308BTP
Canon / Toshiba	PLT-308P
Canon / Toshiba	PLT-604AT
Canon / Toshiba	PLT-704AT
Canon / Toshiba	PLT-704SBT
Canon / Toshiba	PLT-704ST
Canon / Toshiba	PLT-705BT
Canon / Toshiba	PLT-805AT
Canon / Toshiba	PSI-30BX
Canon / Toshiba	PSI-30VX
Canon / Toshiba	PSI-40VX
Canon / Toshiba	PSI-50VX
Canon / Toshiba	PSI-70BT
Canon / Toshiba	PST-25BT
Canon / Toshiba	PST-25ST
Canon / Toshiba	PST-28BT
Canon / Toshiba	PST-30BT

Canon / Toshiba	PST-50BT
Canon / Toshiba	PST-65AT
Canon / Toshiba	PVI-475BT
Canon / Toshiba	PVI-475BX
Canon / Toshiba	PVI-574BX
Canon / Toshiba	PVL-715RST
Canon / Toshiba	PVT-375BT
Canon / Toshiba	PVT-375MV
Canon / Toshiba	PVT-375SC
Canon / Toshiba	PVT-375ST
Canon / Toshiba	PVT-382BT
Canon / Toshiba	PVT-382MV
Canon / Toshiba	PVT-475BT
Canon / Toshiba	PVT-482BT
Canon / Toshiba	PVT-574BT
Canon / Toshiba	PVT-661VT
Canon / Toshiba	PVT-674BT
Canon / Toshiba	PVT-675MV
Canon / Toshiba	PVT-675MVL
Canon / Toshiba	PVT-675MVS
Canon / Toshiba	PVT-681MV
Canon / Toshiba	PVT-681MVL
Canon / Toshiba	PVT-712BT
Canon / Toshiba	PVT-745BTF
Canon / Toshiba	PVT-745BTH
Canon / Toshiba	PVT-770RT
Canon / Toshiba	PVT-781VT
Canon / Toshiba	PVT-781VTE

<https://www.google.com/url?q=https://us.medical.canon/download/ul-xario-200-guidelines-transducers&sa=D&source=hangouts&ust=1581714818335000&usg=AFQjCNGSiMEqjmDh-p5rCU5hHwrUkLEIOQ>



GE

Approved AHP® Disinfectants for Transducers

GE	11L	GE	7S	GE	IC5-9H	GE	P2D-Rse
GE	11L-D	GE	7S-RS	GE	IC9-RS	GE	P6D
GE	12L-RS	GE	8C	GE	L10-22-RS	GE	P6D-RS
GE	12L-SC	GE	8C-RS	GE	L12n-RS	GE	PA2-5P
GE	35-LC	GE	8L	GE	L12n-SC	GE	RAB2-5
GE	3S	GE	8L-RS	GE	L2-9-D	GE	RAB2-5-D
GE	3Sc-RS	GE	9L	GE	L2-9VN-D	GE	RAB2-5L
GE	3S-D	GE	9L-D	GE	L3-12-D	GE	RAB2-5-RS
GE	3Sp	GE	9L-RS	GE	L3-12-RS	GE	RAB2-6-RS
GE	3Sp-D	GE	BE9C	GE	L3-9i-D	GE	RAB4-8-D
GE	3SR	GE	BE9CS	GE	L3-9i-RS	GE	RAB4-8L
GE	3S-RC	GE	BE9CS-D	GE	L4-12t-RS	GE	RAB4-8P
GE	3S-RS	GE	BE9CS-RS	GE	L6-12-RC	GE	RAB4-8-RS
GE	3S-SC	GE	C1-5-D	GE	L6-12-RS	GE	RAB6-D
GE	3V	GE	C1-5-RS	GE	L8-18i-D	GE	RAB6-RS
GE	3V-D	GE	C1-6-D	GE	L8-18i-RS	GE	RIC5-9
GE	4C	GE	C1-6VN-D	GE	L8-18i-SC	GE	RIC5-9A-RS
GE	4C-A	GE	C2-5-RC	GE	M12L	GE	RIC5-9-D
GE	4C-D	GE	C2-7-D	GE	M12L-H	GE	RIC5-9H
GE	4C-RS	GE	C2-7-D-LC	GE	M35	GE	RIC5-9-RS
GE	4C-SC	GE	C2-7VN-D	GE	M4S	GE	RIC5-9W
GE	4D10L	GE	C2-9-D	GE	M4S-D	GE	RIC5-9W-RS
GE	4D16L	GE	C2-9-RS	GE	M4S-RS	GE	RIC6-12-D
GE	4D3C-L	GE	C2-9VN-D	GE	M5Sc-D	GE	RRE5-10-D
GE	4D5C-L	GE	C3-10-D	GE	M5Sc-RS	GE	RSP5-12
GE	4DE7C	GE	C4-8-D	GE	M5S-D	GE	RSP6-12
GE	4S	GE	E8C	GE	M6C	GE	RSP6-16
GE	4Vc-D	GE	E8C-RS	GE	M7C	GE	RSP6-16-D
GE	4V-D	GE	E8CS	GE	M7C-H	GE	RSP6-16-RS
GE	5CS	GE	E8CS-RC	GE	ML6-15	GE	S1-4-D
GE	5S	GE	E8Cs-RS	GE	ML6-15-D	GE	S1-5-D
GE	5S-LC	GE	E8CS-SC	GE	ML6-15-RS	GE	S2-5-D
GE	5Sp	GE	eM6C G2	GE	P2D	GE	Vscan sector
GE	5S-RS	GE	IC5-9-D	GE	P2D-RS		



Mindray

Approved AHP® Disinfectants for Transducers

Mindray	10L24EA
Mindray	2P2
Mindray	2P2P
Mindray	2P2s
Mindray	35C20EA
Mindray	35C50EA
Mindray	35C50EB
Mindray	35C50P
Mindray	3C1
Mindray	3C1p
Mindray	3C1s
Mindray	3C5
Mindray	3C5A
Mindray	3C5P
Mindray	3C5s
Mindray	65C15EA
Mindray	65C15EAV
Mindray	6C2
Mindray	6C2P
Mindray	6C2s
Mindray	75L38EA
Mindray	75L38EB
Mindray	75L38P
Mindray	75L53EA
Mindray	75L60EA
Mindray	7L4
Mindray	7L4A
Mindray	7L4P
Mindray	7L4s
Mindray	7L5
Mindray	7L5P
Mindray	7L5s
Mindray	7L6

Mindray	7L6s
Mindray	C11-3E
Mindray	C11-3s
Mindray	C11-3U
Mindray	C5-1E
Mindray	C5-1s
Mindray	C5-1U
Mindray	C5-2
Mindray	C5-2E
Mindray	C5-2s
Mindray	C6-2E
Mindray	C6-2GE
Mindray	C6-2Gs
Mindray	C6-2GU
Mindray	C7-3E
Mindray	D6-2
Mindray	D6-2A
Mindray	D6-2E
Mindray	D6-2P
Mindray	D7-2
Mindray	D7-2E
Mindray	D7-2s
Mindray	L11-3U
Mindray	L11-4
Mindray	L11-4s
Mindray	L12-3
Mindray	L12-3E
Mindray	L12-4
Mindray	L12-4s
Mindray	L14-6
Mindray	L14-6N
Mindray	L14-6NE
Mindray	L14-6NP

Mindray	L14-6Ns
Mindray	L14-6P
Mindray	L14-6s
Mindray	L14-6WE
Mindray	L14-6Ws
Mindray	L14-6WU
Mindray	L7-3
Mindray	L7-3E
Mindray	L7-3s
Mindray	L9-3U
Mindray	LM14-6E
Mindray	LM14-6s
Mindray	LM16-4U
Mindray	P10-4E
Mindray	P10-4s
Mindray	P10-4U
Mindray	P12-4
Mindray	P12-4s
Mindray	P4-2
Mindray	P4-2E
Mindray	P4-2NE
Mindray	P4-2Ns
Mindray	P4-2s
Mindray	SC5-1E
Mindray	SC5-1U
Mindray	SC6-1s
Mindray	SC6-1U
Mindray	SD8-1E
Mindray	SP5-1E
Mindray	SP5-1s
Mindray	SP5-1U
Mindray	V11-3Hs
Mindray	V11-3HU



Philips

Approved AHP® Disinfectants for Transducers

Philips	S7-2t
Philips	V6-2
Philips	X5-1
Philips	1.9MHz CW Pencil
Philips	1.9MHz TCD
Philips	3D6-2
Philips	3D8-4
Philips	3D9-3v
Philips	3D9-3v)
Philips	5.0MHZ CW Pencil
Philips	7.5MHZ Endo
Philips	BP10-Sec
Philips	C10-3v
Philips	C5-1
Philips	C5-1
Philips	C5-2
Philips	C5-2
Philips	C5-2
Philips	C6-2 Compact
Philips	C6-2 Compact
Philips	C6-3
Philips	C8-4v
Philips	C8-4v
Philips	C8-5
Philips	C8-5 Compact
Philips	C9-2
Philips	C9-3io
Philips	C9-3v
Philips	C9-4
Philips	C9-4ec
Philips	C9-4v

Philips	C9-4v Compact
Philips	C9-5 Compact
Philips	C9-5ec
Philips	D2cwc
Philips	D2tcd
Philips	D5cwc
Philips	eL18-4
Philips	L10-4lap
Philips	L11-3
Philips	L12-3
Philips	L12-3
Philips	L12-4
Philips	L12-4
Philips	L12-4 Compact
Philips	L12-4 Compact
Philips	L12-5 38mm
Philips	L12-5 50 mm
Philips	L12-5 50mm Compact
Philips	L12-5 50mm Compact
Philips	L15-7io
Philips	L15-7io
Philips	L15-7io
Philips	L17-5
Philips	L18-5
Philips	L8-4
Philips	L9-3
Philips	OMNI III TEE
Philips	S12
Philips	S12-4
Philips	S12-4 Compact
Philips	S12-4 Compact

Philips	S3-1
Philips	S4-1
Philips	S4-1
Philips	s4-1
Philips	S4-2
Philips	S4-2 Compact
Philips	S4-2 Compact
Philips	S5-1
Philips	S5-1
Philips	S5-2
Philips	S6-2mpt
Philips	S7-3t
Philips	S8
Philips	S8-3
Philips	S8-3 Compact
Philips	S8-3 Compact
Philips	S8-3t
Philips	S9-2
Philips	TCD
Philips	V6-2
Philips	V8-4
Philips	VL13-5
Philips	X3-1
Philips	X5-1
Philips	X6-1
Philips	X7-2
Philips	X7-2t
Philips	X8-2t
Philips	X8-2ti



Siemens

Approved AHP® Disinfectants for Transducers

Siemens	11M3	Siemens	6C2	Siemens	8V5	Siemens	L10-5v
Siemens	15L8	Siemens	6L3	Siemens	C5-2v	Siemens	LA435
Siemens	2MHxCW	Siemens	7.5L40A	Siemens	CA123	Siemens	LA522E
Siemens	3V2c	Siemens	7V3c	Siemens	CA431	Siemens	LA523
Siemens	4C1	Siemens	8C4	Siemens	EC10-5W	Siemens	PA023E
Siemens	4V1	Siemens	8L5	Siemens	EC-10C5	Siemens	PA122E
Siemens	4V1c	Siemens	8L5T	Siemens	EC1123	Siemens	PA230E
Siemens	5.0C50A	Siemens	8V3	Siemens	EC9-4	Siemens	TV-A
Siemens	5MHz CW	Siemens	8V3	Siemens	EV10-5		
Siemens	5V2c	Siemens	8V3c	Siemens	EV-8C4		

<https://www.siemens-healthineers.com/en-us/ultrasound/ultrasound-transducer-catalog>



Sonosite

Approved AHP® Disinfectants for Transducers

Sonosite	C60xp	Sonosite	HFL50x	Sonosite	ICTxp	Sonosite	L52x
Sonosite	HFL38xp	Sonosite	HFL50xp	Sonosite	L52n		

<https://www.sonosite.com/support/cleaners-disinfectants>

*Diversey is currently working with Fuji Sonosite to execute additional testing on Sonosite Transducers for broader approval of Oxivir on their devices.



Approved AHP® Disinfectants for Transducers